

Test results

Virus Types	(NO)	$\lg(V_{a0h})$ ($\lg\text{TCID}_{50}/\text{mL}$)	$\lg(V_{b2h})$ ($\lg\text{TCID}_{50}/\text{mL}$)	$\lg(V_{c2h})$ ($\lg\text{TCID}_{50}/\text{mL}$)
SARS-CoV-2 MDCK	1	6.67	6.53	4.32
	2	6.63	6.51	4.35
	3	6.59	6.48	4.27
Average Value of $\lg\text{TCID}_{50}/\text{mL}$		6.63	6.51	4.31
Antiviral Activity Value		2.19		
Antiviral Activity Rate (%)		99.36		



End

